

TES-1360A Humidity/Temperature Meter



APPLICATIONS

Laboratory
Quality Control
Air-Conditioning
Process Engineering



SPECIFICATION

| | |
|------------------------|-------------------------------------|
| Range: Humidity | 10% to 95% R.H. |
| Temperature | -20 °C to +60 °C ; -4 °F to +140 °F |
| Resolution | 0.1% R.H. 0.1 °C , 0.1 °F |

Accuracy (after calibration)

| | |
|--------------------|--|
| Humidity | +/- 3% R.H. (at 25 °C , 30...95% R.H.) |
| Temperature | +/- 0.8 °C , +/- 1.5 °F |

Response time

| | |
|--------------------|---|
| Humidity | 45% R.H. --> 95% R.H. <= 3min 95% R.H. --> 45% R.H. <=5min |
| Temperature | 1 °C /2 sec |
| Accessories | Carrying case, Instruction manual, Battery. |

| | |
|------------------------------------|--|
| Operating/Storage Condition | 0 °C to 50 °C < 80% R.H. -10 °C to 60 °C < 70% R.H. |
| Power Source | 10mV DC/1 °C , 10mV DC/1 °F |
| Size | 270mm(L)x68mm(W)x25mm(H) |
| Weight | 200g |
| Accessories | Instruction Manual, Battery and Carrying Case |